

Real-Time Water Quality Deployment Report

Lower Churchill River Network

July 24 to Aug 22, 2010



Government of Newfoundland & Labrador
Department of Environment and Conservation
Water Resources Management Division

General

- Department of Environment and Conservation staff monitors the real-time web pages regularly.
- This monthly deployment period illustrates and discusses water quality related events from July 24 to August 22, 2010; a period of 29 days.
- On July 23, 2010, real-time water quality monitoring instruments were deployed at all four Lower Churchill River Stations.

Maintenance and Calibration of Instrument

- As part of the Quality Assurance and Quality Control protocol (QAQC), an assessment of the reliability of data recorded by an instrument is made at the beginning and end of the deployment period. The procedure is based on the approach used by the United States Geological Survey.
 - ▶ At deployment and removal, a QA/QC Sonde is temporarily deployed along side the Field Sonde. Values for temperature, pH, conductivity, dissolved oxygen and turbidity are compared between the two instruments. Based on the degree of difference between parameters recorded by the Field Sonde and QAQC Sonde at deployment, a qualitative statement is made on the data quality (Table 1).
 - ▶ At the end of a deployment period, readings are taken in the water body from the Field Sonde before and after a thorough cleaning in order to assess the degree of biofouling. During calibration in the laboratory, an assessment of calibration drift is made and the two error values are combined to give Total Error (T_e). If T_e exceeds a predetermined data correction criterion, a correction based on T_e is applied to the dataset using linear interpolation.

Table 1: Ranking classifications for deployment and removal

Parameter	Rank				
	Excellent	Good	Fair	Marginal	Poor
Temperature (oC)	<=+-0.2	>+-0.2 to 0.5	>+-0.5 to 0.8	>+-0.8 to 1	<+-1
pH (unit)	<=+-0.2	>+-0.2 to 0.5	>+-0.5 to 0.8	>+-0.8 to 1	>+-1
Sp. Conductance (μ S/cm)	<=+-3	>+-3 to 10	>+-10 to 15	>+-15 to 20	>+-20
Sp. Conductance > 35 μ S/cm (%)	<=+-3	>+-3 to 10	>+-10 to 15	>+-15 to 20	>+-20
Dissolved Oxygen (mg/L) (% Sat)	<=+-0.3	>+-0.3 to 0.5	>+-0.5 to 0.8	>+-0.8 to 1	>+-1
Turbidity <40 NTU (NTU)	<=+-2	>+-2 to 5	>+-5 to 8	>+-8 to 10	>+-10
Turbidity > 40 NTU (%)	<=+-5	>+-5 to 10	>+-10 to 15	>+-15 to 20	>+-20

- It should be noted that the temperature sensor on any sonde is the most important. All other parameters can be broken down into three groups: temperature dependant, temperature compensated and temperature independent. Because the temperature sensor is not isolated from the rest of the sonde the entire sonde must be at the same temperature before the sensor will stabilize. The values may take some time to climb to the appropriate reading; if a reading is taken too soon it may not accurately portray the water body.
- Deployment and removal comparison rankings for the Churchill River stations deployed between July 23 and August 22, 2010 are summarized in Table 2.

Table 2: Comparison rankings for Churchill River stations, July 23 – August 22, 2010

Station Churchill River	Date	Action	Comparison Ranking				
			Temperatur e	pH	Conductivity	Dissolved Oxygen	Turbidity
Below Muskrat Falls	July 23, 2010	Deployment	Excellent	Good	Excellent	Good	Excellent
	Aug 22, 2010	Removal	Good	Good	Excellent	Excellent	Excellent
Above Muskrat Falls	July 23, 2010	Deployment	Good	Excellent	Excellent	Excellent	Excellent
	Aug 22, 2010	Removal	Good	Good	Excellent	Excellent	Excellent
Below Grizzle Rapids	July 23, 2010	Deployment	Excellent	Excellent	Excellent	Excellent	Excellent
	Aug 22, 2010	Removal	Excellent	Good	Excellent	Excellent	Excellent
Below Metchin River	July 23, 2010	Deployment	Excellent	Good	Excellent	Excellent	Excellent
	Aug 22, 2010	Removal	Excellent	Good	Good	Excellent	Excellent

- A transmission error at the station below Grizzle Rapids prevented data from being transmitted in real time between July 12 and August 22. The instruments internal log file was recovered to provide data for this period.

Data Interpretation

- The following graphs and discussion illustrate significant water quality-related events from July 24 to August 22 in the Lower Churchill River Network.
- With the exception of water quantity data (stage), all data used in the preparation of the graphs and subsequent discussion below adhere to this stringent QAQC protocol. Water Survey of Canada is responsible for QAQC of water quantity data. Corrected data can be obtained upon request. Where appropriate, corrected data for water quality parameters are indicated.

Churchill River below Muskrat Falls

- Water temperature ranged from 15.8 to 19.00°C during this deployment period (Figure 1).
- Water temperature fluctuates slightly throughout the deployment period but generally remains consistent. This trend is expected during the peak summer air temperatures (Appendix 1). Water temperature fluctuates diurnally.




Figure 1: Water temperature at Churchill River below Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- pH ranges between 6.98 and 7.27 pH units and are consistent throughout the deployment period (Figure 2).
- All values during the deployment are within the recommended range as suggested by the CCME Guidelines for the Protection of Aquatic Life (indicated in blue on Figure 2).




Figure 2: pH at Churchill River below Muskrat Falls

- Specific conductivity ranged from 17.7 to 20.1 μ S/cm during the deployment period (Figure 3). Specific conductance generally increases throughout the deployment period.
- Stage is included in Figure 3 to illustrate the inverse relationship between conductivity and water level. Stage generally remains stable throughout the deployment period with slight increases and decreases. As stage increases, specific conductivity decreases (indicated by red arrows on Figure 13).




Figure 3: Specific conductivity at Churchill River below Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- The saturation of dissolved oxygen ranged from 106.2 to 116.0% and a range of 9.96 to 10.98mg/l was found in the concentration of dissolved oxygen with a median value of 10.66 mg/l (Figure 4).
- All values were above both the minimum CCME Guideline for the Protection of Other Life Stage Cold Water Biota of 6.5 mg/l and the minimum CCME Guideline for the Protection of Early Life Stage Cold Water Biota value of 9.5 mg/l. The guidelines are indicated in green on Figure 4
- Dissolved oxygen content remains relatively consistent throughout the deployment period. This trend is expected given the stable and warm water temperatures during the mid-summer (Figure 1).




Figure 4: Dissolved oxygen at Churchill River below Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- A range of 0.9 to 25.7 NTU was recorded for turbidity for this deployment period (Figure 5). A median value of 3.1 NTU indicates there is a consistent natural background turbidity value at this station.
- One significant spike (to 261.5NTU) was recorded on August 9 and corresponds with a small (but rare) rainfall event (Appendix 1).




Figure 5: Turbidity at Churchill River below Muskrat Falls

Churchill River above Muskrat Falls

- Water temperature ranges from 15.32 to 19.63°C during this deployment period (Figure 6).
- Water temperature is stable throughout the mid-summer deployment period. This trend is expected given the warm ambient air temperature (Appendix 1). Water temperature fluctuates diurnally.




Figure 6: Water temperature at Churchill River above Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- pH ranges between 7.12 and 7.32 pH units (Figure 7). pH values are consistent throughout the deployment period.
- All values are within the recommended range for pH as suggested by the CCME Guidelines for the Protection of Aquatic Life (indicated in blue on Figure 7).




Figure 7: pH at Churchill River above Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- Specific conductivity ranges between 16.1 and 18.0 μ S/cm during the deployment period (Figure 8).
- Stage is included in Figure 8 to illustrate the relationship between conductivity and water level. Stage, generally decreases throughout the deployment period and conductivity increases. There are some instances where stage is increasing which correspond with decreases in specific conductivity (indicated by red arrows on Figure 8).




Figure 8: Specific conductivity at Churchill River above Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- The saturation of dissolved oxygen ranged from 93.2 to 100.7.8% and a range of 8.92 to 9.45 mg/l was found in the concentration of dissolved oxygen with a median value of 9.26 mg/l (Figure 9).
- All values were above the minimum CCME Guideline for the Protection of Other Life Stage Cold Water Biota of 6.5 mg/l. All values were below the minimum CCME Guideline for the Protection of Early Life Stage Cold Water Biota value of 9.5 mg/l. The guidelines are indicated in green on Figure 9.
- Dissolved oxygen content is stable throughout the deployment period and is relatively low compared to other deployment periods. This trend is partly expected given the warm, consistent water temperatures during the deployment period (Figure 6).




Figure 9: Dissolved oxygen at Churchill River above Muskrat Falls

Lower Churchill River, Newfoundland and Labrador

- A range of 1.0 to 431.0 NTU was recorded for turbidity for this deployment period (Figure 10). A median value of 2.0 NTU indicates there is a consistent natural background turbidity value at this station.
- Turbidity values were generally very stable for this station throughout the deployment period. Most of the values recorded were within 1.0 to 16.0 NTU. One event, near the end of the deployment period, recorded spikes up to 431.0 NTU. This event does not correspond with any precipitation or weather related events (Appendix 1) however it was during a time of very low water level. The exact cause of this event is unknown.
- Other smaller events (>15NTU) were recorded and all correspond with light to moderate rainfall events (Appendix 1).




Figure 10: Turbidity at Churchill River above Muskrat Falls

Churchill River below Grizzle Rapids

- A transmission error occurred on July 12 and the instruments internal log file was recovered to provide information for this period. Stage and flow data is not available for this time.
- Water temperature ranges from 15.30 to 18.95°C during this deployment period (Figure 11).
- Water temperature is relatively stable throughout the deployment period (Figure 11). This trend is expected given the continuing warm summer air temperatures (Appendix 1). Water temperature fluctuates diurnally.




Figure 11: Water temperature at Churchill River below Grizzle Rapids

Lower Churchill River, Newfoundland and Labrador

- pH ranges between 6.90 and 7.30 pH units (Figure 12). pH values are consistent throughout the deployment period.
- All values during the deployment are within the recommended range as suggested by the CCME Guidelines for the Protection of Aquatic Life (indicated in blue on Figure 12).




Figure 12: pH at Churchill River below Grizzle Rapids

- Data recovered from the instruments internal log file after July 12 only measures conductivity to 0 decimal places.
- Specific conductivity ranges between 18.0 and 20.0 μ S/cm during the deployment period (Figure 13). Specific conductance generally increases throughout the deployment period.




Figure 13: Specific conductivity at Churchill River below Grizzle Rapids

Lower Churchill River, Newfoundland and Labrador

- The saturation of dissolved oxygen ranged from 92.4 to 99.7% and a range of 8.91 to 9.47 mg/l was found in the concentration of dissolved oxygen with a median value of 9.19 mg/l (Figure 14).
- All values were above the minimum CCME Guideline for the Protection of Other Life Stage Cold Water Biota of 6.5 mg/l. All values were below the minimum CCME Guideline for the Protection of Early Life Stage Cold Water Biota value of 9.5 mg/l. The guidelines are indicated in green on Figure 14.
- Dissolved oxygen content is stable throughout the deployment period and is relatively low compared to other deployment periods. This trend is partly expected given the warm, consistent water temperatures during the deployment period (Figure 11).




Figure 14: Dissolved oxygen at Churchill River below Grizzle Rapids

Lower Churchill River, Newfoundland and Labrador

- A range of 0.0 to 1.6NTU was recorded for turbidity for this deployment period (Figure 15). A median value of 0.0 NTU indicates this site is naturally clear with out significant background turbidity.
- Turbidity remains at 0.0NTU for the entire deployment period with one hourly reading at 1.6NTU.




Figure 15: Turbidity at Churchill River below Grizzle Rapids

Churchill River below Metchin River

- Water temperature ranges from 15.70 to 18.10°C during this deployment period (Figure 16).
- Water temperature is stable throughout the mid-summer deployment period. This trend is expected given the warm ambient air temperature (Appendix 1). Water temperature fluctuates diurnally.




Figure 16: Water temperature at Churchill River below Metchin River

Lower Churchill River, Newfoundland and Labrador

- pH ranges between 6.96 and 7.33 pH units (Figure 17). pH values are consistent throughout the deployment period.
- All values during the deployment are within the recommended range as suggested by the CCME Guidelines for the Protection of Aquatic Life (indicated in blue on Figure 17).




Figure 17: pH at Churchill River below Metchin River

Lower Churchill River, Newfoundland and Labrador

- Specific conductivity ranged from 16.0 to 22.4 μ S/cm during the deployment period (Figure 18). Specific conductance fluctuates slightly throughout the deployment period and shows a general increasing trend.
- Variability in specific conductance most often can be related back to weather events. The nearest weather station to the site below Metchin River is located in Churchill Falls. However, for much of the time between May and November 2010, precipitation data is missing from the dataset. Generally, as stage increases, conductivity decreases. This trend is indicated by the red arrows on Figure 18.




Figure 18: Specific conductivity at Churchill River below Metchin River

Lower Churchill River, Newfoundland and Labrador

- The saturation of dissolved oxygen ranged from 91.5 to 98.6% and a range of 8.88 to 9.45 mg/l was found in the concentration of dissolved oxygen with a median value of 9.13mg/l (Figure 19).
- All values were above the minimum CCME Guideline for the Protection of Other Life Stage Cold Water Biota of 6.5 mg/l. All values were below the minimum CCME Guideline for the Protection of Early Life Stage Cold Water Biota value of 9.5 mg/l.
- Dissolved oxygen content is stable throughout the deployment period and is relatively low compared to other deployment periods. This trend is partly expected given the warm, consistent water temperatures during the deployment period (Figure 16).




Figure 19: Dissolved oxygen at Churchill River below Metchin River

Lower Churchill River, Newfoundland and Labrador

- A range of 0.0 to 3.4 NTU was recorded for turbidity for this deployment period (Figure 20). A median value of 0.0 NTU indicates this site is naturally clear with out significant background turbidity values.
- Turbidity values remain at or near 0.0NTU for the entire deployment period. Spikes are of minimal magnitude (<3.5NTU).




Figure 20: Turbidity at Churchill River below Metchin River


Conclusions

- Instruments at water quality monitoring stations on the Lower Churchill River were deployed at all four stations between July 24 and August 22.
- No significant water quality events were captured during this time. In most cases, weather related events or increase/decreases in water level could be used to explain the fluctuations.
- All values recorded were within ranges as suggested by the CCME Guidelines for the Protection of Aquatic Life for pH. Dissolved oxygen content was below the CCME Guideline for the Protection of Aquatic Life at Early Life Stages at all stations except for the station below Muskrat Falls.
- A transmission error at the station below Grizzle Rapids prevented data from being transmitted in real time between July 12 and August 22. The instruments internal log file was recovered to provide data for this period.

Prepared by:
Grace Gillis
Department of Environment and Conservation
Water Resources Management Division
Phone: 709.896.5542
Fax: 709.896.9566

Appendix 1

**Mean Daily Air Temperature and Total Precipitation
Goose Bay, July 24 - August 22, 2010**



Mean Daily Air Temperature and Total Precipitation
Churchill Falls, July 24 - August 22, 2010

